

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

<b>1. Department:</b> Seaside Heights Police Department 116 Sherman Ave. Seaside Heights, NJ 08751	<b>2. Contact:</b> Tpr. I Steve Dellanoce #6027  <b>3. Phone Number:</b> 609-947-9211	<b>4. Date:</b> 03/21/2012  <b>5. County:</b> Ocean
<b>6. Alcotest Instrument Serial Number:</b>  ARWJ-0018	<b>7. Simulator Component Serial Number:</b>  N/A	<b>8. Temperature Probe Component Serial Number:</b>  N/A

**9. Reason for Service:**  
Alcotest instrument returned from outside maintenance and placed back in service.

**10. Comments:**  
See Draeger Return and Repair form.

- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☐ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Steve Dellanoce #6027  
Name & Badge Number (Print)

Tpr. I Steve Dellanoce #6027 3/21/12  
Signature Date



# RETURN AND REPAIR FORM

# Dräger

Customer Information:

B:

S:

Company Name: Seaside Hgts. P.D., Seaside Heights NJ

Date Received: 03/06/12 Date given to service: 03/06/12

Carrier: ☒ FedEx ☐ UPS ☐ USPS Shipping Method: ☐ GRD ☒ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☒ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole  
 Serial Number: AR WJ-0018  
 Printer Serial#: AR  
 Sim Serial#: DD  
 Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
6808486	Plate	1	N/C W
6808455	Fuel cell	1	N/C W
mp cal 71	Calibration	1	N/C W
mp labor	labor	1.5	N/C W

Repair Notes: Replaced Fuel cell # ARXD-1487 Fuel cell was  
busted, installed a new fuel cell # ARCK-0917.

Cal w/DC & ops check

Service Technician: [Signature]

Date: 3-16-12



# Packing Slip

Customer no      Order No      Order date  
150060725    10536805    03/16/2012

Packing slip no.      Ship date  
80671804      03/16/2012

Please reference on inquiries

ship to

150060729  
BOROUGH OF SEASIDE HEIGHTS POLICE  
ATTN: CHIEF THOMAS J. BOYD  
OCEAN COUNTY  
116 SHERMAN AVENUE  
SEASIDE HEIGHTS, NJ 08751  
USA

bill to

Your Purchase Order Number

7110: ARWJ-0018  
Any questions? Please contact:  
MARA HERRERA

Your sales person

Don Pouliot  
Phone: 973-398-3228

Customer  
BOROUGH OF SEASIDE HEIGHTS  
MUNICIPAL BUILDING  
OCEAN COUNTY  
901 BLVD-PO BOX 38  
SEASIDE HEIGHTS, NJ 08751  
USA

Delivery terms

FA  
FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110: ARWJ-0018
0010	1 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA

Draeger Safety Diagnostics Inc.  
Accounting Address:  
101 Technology Drive  
Pittsburgh, PA 15275  
Tel: (412) 788-5537  
Fax: (412) 788-5598

Remit Wire Transfers To:  
Citizens Bank  
Acct. # 6209426615  
Acct. Name: Draeger Safety Diagnostics Inc.  
Transit Number: 036076150  
Federal ID Number: 84-1600159

REMIT TO:  
Draeger Safety Diagnostics Inc.  
P.O. Box 200337  
Pittsburgh, PA 15251-0337



# Packing Slip

Customer no      Order No      Order date  
150060725    10536805    03/16/2012

Packing slip no.      Ship date  
80671804      03/16/2012

Please reference on inquiries

Page      2 / 2

Item#	sh. Quant.	Part no. Description			
0030	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0040	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			



# Dräger safety

Internal Use

## Return & Repair Form

### Order Information:

Order # Seaside Hsts PD Cust # 10725/10729  
 Return: 7/6/07 Carrier: Fed Ex Method: Priority Overnight

### Unit Information:

Unit: 7110 7410 5510 6810 Serial # ARWJ-0018  
 Options: A B PLUS SCREENER Printer Serial # \_\_\_\_\_  
 Whole Inst. Top Half Simulator Serial # \_\_\_\_\_  
 Probe Serial # \_\_\_\_\_  
 Unit ☐ Returned to stock on: \_\_\_\_\_

### Accessories: (check all that apply):

☒ 110VAC Adapter ☐ Regulator ☐ Mag Card Rdr # \_\_\_\_\_  
☐ Printer Paper ☐ Printer Ribbon ☐ Casio # \_\_\_\_\_  
☐ Mounting pieces ☐ Carrying Case ☐ Dry Gas \_\_\_\_\_  
 Other (specify): \_\_\_\_\_

Expiry Date: \_\_\_\_\_

### Test Information:

Test # \_\_\_\_\_

Reason for Return: LOW FC RESULTS

Item #	Description	Qty	Total
6808455	FC	1	WARR
MPCK #1	7110 CAC	1	WARR
MPLABOR	LABOR	5	WARR

Notes: REPLACED F/C / CAC w/ QC OPS CHECKS

Signature: \_\_\_\_\_

Date: 09.06.07



STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

<b>1. Department:</b> Seaside Heights Police Department 116 Sherman Ave. Seaside Heights, NJ 08751	<b>2. Contact:</b> <b>Tpr. I Steve Dellanoce #6027</b>	<b>4. Date:</b> <b>03/01/2012</b>
	<b>3. Phone Number:</b> <b>609-947-9211</b>	<b>5. County:</b> <b>Ocean</b>
<b>6. Alcotest Instrument Serial Number:</b> <b>ARWJ-0018</b>	<b>7. Simulator Component Serial Number:</b> <b>N/A</b>	<b>8. Temperature Probe Component Serial Number:</b> <b>N/A</b>

**9. Reason for Service:**

While reviewing the performance history of the instrument prior to recalibration low EC results on some control tests were noted.

**10. Comments:**

Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Please replace Fuel Cell.

- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☒ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Steve Dellanoce #6027  
Name & Badge Number (Print)

Tpr I [Signature] #6027  
Signature

3/1/12  
Date