

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01108	Calib. Date: 02/24/2015
Certification File No.:	01109	Cert. No.: 00025
Linearity File No.:	01110	Cert. No.: 00018
Solution File No.:	01159	Lin. No.: 00018
Sequential File No.:	01159	Soln. No.: 00208
		File Date: 06/21/2015
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 0757

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:19D		
Control 1 EC	0.101%	10:20D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	10:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:21D		
Control 2 EC	0.101%	10:21D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:22D		
Control 3 EC	0.101%	10:23D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	10:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 06/21/2015