

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01187	Calib. Date: 08/03/2015
Certification File No.:	01188	Calib. No.: 00026
Linearity File No.:	01189	Cert. Date: 08/03/2015
Solution File No.:	01206	Cert. No.: 00019
Sequential File No.:	01206	Lin. Date: 08/03/2015
		Lin. No.: 00019
		Soln. Date: 08/17/2015
		Soln. No.: 00216
		File Date: 08/17/2015

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWL S3-0440
Control Solution %: 0.100%		Expires: 06/20/2016
Solution Control Lot: 14F130		Bottle No.: 0758

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:46D		
Control 1 EC	0.096%	00:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:47D		
Control 2 EC	0.096%	00:48D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:48D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:48D		
Control 3 EC	0.097%	00:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HEFFERNAN

First Name: JUSTIN

MI: R

Signature: _____



Badge No.: 9028

Date: 08/17/2015