

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01261	Calib. Date: 03/01/2016
Certification File No.:	01262	Calib. No.: 00029
Linearity File No.:	01263	Cert. No.: 00021
Solution File No.:	01297	Lin. Date: 03/01/2016
Sequential File No.:	01297	Lin. No.: 00021
		Soln. No.: 00231
		File Date: 05/01/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWL S3-0441
Control Solution %:	0.100%		Expires: 06/13/2017
Solution Control Lot:	15F139		Bottle No.: 0720

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:11D		
Control 1 EC	0.097%	05:11D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	05:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:12D		
Control 2 EC	0.096%	05:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	05:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:13D		
Control 3 EC	0.096%	05:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	05:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:15D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 05/01/2016