

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT


1. Department: Seaside Heights Police Department 116 Sherman Ave Seaside Heights, NJ 08751	2. Contact: Sgt. Rich Roemmele 3. Phone Number: 732-793-1800	4. Date: 10/28/2016 5. County: Ocean
6. Alcotest Instrument Serial Number: ARWJ-0018	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A

9. Reason for Service:

The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:

See Draeger Return and Repair form.



Last Known Sequential File #: N/A

☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.

☐ 12. The above Instrument/Component is placed out of service pending further evaluation.

☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704
Name & Badge Number (Print)

Tpr. I K W. Alcott #6704 **10/28/16**
Signature Date

RETURN AND REPAIR FORM

Dräger

Customer Information

B:

S:

Company Name: Seaside Heights P.D., Seaside Heights, NJ

Date Received: 09/15/2016

Date given to service: 09/15/2016

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method:

☐ GRD
☒ NDA-PRI

☐ 3DAY ☐ 2DAY
☐ NDA-STD

Product: ☐ 6510 ☐ 6810
☒ 7110 ☐ 8610
☐ 7510 ☐ DT5000
☐ 7410 Upper-half ☐ 9510
☐ 7410 Whole

Serial Number: AR WJ-0018

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: _____

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

Accessories

☐ 110V A/C Adapter
☐ Mouthpieces
☐ Dry Gas

☐ Regulator
☐ 9510 Stylus
☐ Other (please specify) _____

☐ Printer Ribbon
☐ 9510 Top Cover

☐ Printer Paper
☐ Carrying Case

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	

Repair Notes: Recalibrated unit, performed and passed all QC.

CAL W/QC AND OPS CHECK

Service Technician: BC

Date: 10/12/2016

Packing Slip

Customer no Order No Order date
150060725 10855301 10/12/2016

Packing slip no. Ship date
81032757 10/12/2016
Please reference on inquiries

Ship to

150060729
BOROUGH OF SEASIDE HEIGHTS POLICE
ATTN: SGT RICH ROEMMELE
OCEAN COUNTY
116 SHERMAN AVENUE
SEASIDE HEIGHTS, NJ 08751-2055
USA

Payer

Your Purchase Order Number

10/12/2016
SGT RICH ROEMMELE
Any questions? Please contact:
ROSEMARIE HUGHES

Customer

BOROUGH OF SEASIDE HEIGHTS
MUNICIPAL BUILDING
OCEAN COUNTY
901 BLVD-PO BOX 38
SEASIDE HEIGHTS, NJ 08751-0038
USA

Your sales person

CHRISTOPHER FAIRCHILD
Phone: 866-385-5900
Fax: 972-929-1260

Delivery terms

PPA
PREPAY AND ADD

Page 1 / 2

Number of packages: 1

Item#	sh. Quant.	Part no. Description			
		Waybill number: 1001891751960000875100677251656735 Ship via: Fed Ex P1			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARWJ-0018 REC'D 9/14/2016			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 1 / 0 EA			

Draeger Safety Diagnostics Inc.
Accounting Address:
101 Technology Drive
Pittsburgh, PA 15275
Tel: (412) 788-5537
Fax: (412) 788-5598

Remit Wire Transfers To:
Citizens Bank
Acct. # 6209426615
Acct. Name: Draeger Safety Diagnostics Inc.
Transit Number: 036076150
Federal ID Number: 84-1600159

REMIT TO:
Draeger Safety Diagnostics Inc.
P.O. Box 536410
Pittsburgh, PA 15253-5906

Packing Slip

Customer no	Order No	Order date
150060725	10855301	10/12/2016

Packing slip no.	Ship date
81032757	10/12/2016
Please reference on inquiries	

Page 2 / 2

Item#	sh. Quant.	Part no. Description			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.5 / 0.0 EA			
0030	1 EA	MPINS INSURANCE AND HANDLING CHARGE Ordered / Back ordered : 1 / 0 EA			