

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	Calib. No.:	00033
Calibration File No.:	01427	Calib. Date:	04/21/2017
Certification File No.:	01428	Cert. Date:	04/21/2017
Linearity File No.:	01429	Lin. Date:	04/21/2017
Solution File No.:	01492	Soln. Date:	07/19/2017
Sequential File No.:	01492	File Date:	07/19/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0441
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1051

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:09D		
Control 1 EC	0.098%	21:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10D		
Control 2 EC	0.097%	21:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	21:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:12D		
Control 3 EC	0.097%	21:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	21:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____



Badge No.: 120

Date: 07/19/2017