

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01231 Calib. Date: 11/10/2015 Calib. No.: 00028
Certification File No.: 01232 Cert. Date: 11/10/2015 Cert. No.: 00020
Linearity File No.: 01233 Lin. Date: 11/10/2015 Lin. No.: 00020
Solution File No.: 01252 Soln. Date: 02/06/2016 Soln. No.: 00225
Sequential File No.: 01252 File Date: 02/06/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 06/13/2017
Solution Control Lot: 15F139 Bottle No.: 0500

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:40S		
Control 1 EC	0.099%	07:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		
Control 2 EC	0.098%	07:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		
Control 3 EC	0.098%	07:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN.

MI: M

Signature: _____



Badge No.: 120

Date: 02/06/2016