

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01689	Calib. Date: 01/15/2019
Certification File No.:	01690	Calib. No.: 00038
Linearity File No.:	01691	Cert. Date: 01/15/2019
Solution File No.:	01711	Cert. No.: 00029
Sequential File No.:	01711	Lin. Date: 01/15/2019
		Lin. No.: 00029
		Soln. Date: 03/14/2019
		Soln. No.: 00290
		File Date: 03/14/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0429

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:13S		
Control 1 EC	0.101%	23:14S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:14S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:15S		
Control 2 EC	0.100%	23:15S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	23:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:16S		
Control 3 EC	0.100%	23:17S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	23:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____



Badge No.: 120

Date: 03/14/2019