

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01775	Calib. Date: 11/27/2019
Certification File No.:	01776	Calib. No.: 00040
Linearity File No.:	01777	Cert. Date: 11/27/2019
Solution File No.:	01783	Cert. No.: 00031
Sequential File No.:	01783	Lin. Date: 11/27/2019
		Lin. No.: 00031
		Soln. Date: 12/20/2019
		Soln. No.: 00301
		File Date: 12/20/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWL S3-0440
Control Solution %: 0.100%		Expires: 01/08/2021
Solution Control Lot: 19020		Bottle No.: 1485

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:26S		
Control 1 EC	0.097%	04:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:27S		
Control 2 EC	0.095%	04:28S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:28S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:29S		
Control 3 EC	0.095%	04:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	04:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: *Shawn Heckler #120*

Badge No.: 120

Date: 12/20/2019