

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01875 Calib. Date: 03/16/2021 Calib. No.: 00043
Certification File No.: 01876 Cert. Date: 03/16/2021 Cert. No.: 00034
Linearity File No.: 01877 Lin. Date: 03/16/2021 Lin. No.: 00034
Solution File No.: 01880 Soln. Date: 04/13/2021 Soln. No.: 00328
Sequential File No.: 01880 File Date: 04/13/2021

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 01/13/2023
Solution Control Lot: 21010 Bottle No.: 0557

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:24D		
Control 1 EC	0.101%	13:25D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:25D		
Control 2 EC	0.100%	13:26D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.101%	13:26D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:27D		
Control 3 EC	0.100%	13:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:28D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 04/13/2021