

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01875 **Calib. Date:** 03/16/2021 **Calib. No.:** 00043
Certification File No.: 01876 **Cert. Date:** 03/16/2021 **Cert. No.:** 00034
Linearity File No.: 01877 **Lin. Date:** 03/16/2021 **Lin. No.:** 00034
Solution File No.: 01894 **Soln. Date:** 05/30/2021 **Soln. No.:** 00330
Sequential File No.: 01894 **File Date:** 05/30/2021

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0441
Control Solution %: 0.100% **Expires:** 01/13/2023
Solution Control Lot: 21010 **Bottle No.:** 0499

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:55D		
Control 1 EC	0.100%	09:55D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	09:55D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:56D		
Control 2 EC	0.099%	09:57D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:58D		
Control 3 EC	0.100%	09:58D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:59D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/30/2021