

# Calibrating Unit New Standard Solution Report

<b>Equipment:</b> Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018	
<b>Location:</b> SEASIDE HEIGHTS POLICE		
Calibration File No.: 01975	Calib. Date: 01/20/2022	Calib. No.: 00045
Certification File No.: 01976	Cert. Date: 01/20/2022	Cert. No.: 00036
Linearity File No.: 01977	Lin. Date: 01/20/2022	Lin. No.: 00036
Solution File No.: 01985	Soln. Date: 02/12/2022	Soln. No.: 00343
Sequential File No.: 01985	File Date: 02/12/2022	
<b>Calibrating Unit:</b> WET	<b>Model No.:</b> CU-34	<b>Serial No.:</b> DDWL S3-0441
<b>Control Solution %:</b> 0.100%		<b>Expires:</b> 01/13/2023
<b>Solution Control Lot:</b> 21010		<b>Bottle No.:</b> 0054

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:34S		
Control 1 EC	0.098%	16:34S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.099%	16:34S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:35S		
Control 2 EC	0.097%	16:36S	34.1°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:36S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:37S		
Control 3 EC	0.097%	16:37S	34.1°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:37S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:38S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: -

Signature: \_\_\_\_\_

Badge No.: 128

Date: 02/12/2022

