

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01975 **Calib. Date:** 01/20/2022 **Calib. No.:** 00045
Certification File No.: 01976 **Cert. Date:** 01/20/2022 **Cert. No.:** 00036
Linearity File No.: 01977 **Lin. Date:** 01/20/2022 **Lin. No.:** 00036
Solution File No.: 02066 **Soln. Date:** 07/04/2022 **Soln. No.:** 00351
Sequential File No.: 02066 **File Date:** 07/04/2022

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0441
Control Solution %: 0.100% **Expires:** 12/28/2023
Solution Control Lot: 21480 **Bottle No.:** 0244

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:20D		
Control 1 EC	0.099%	22:21D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:21D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:22D		
Control 2 EC	0.097%	22:22D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		
Control 3 EC	0.097%	22:24D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 07/04/2022