

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 02115 Calib. Date: 12/09/2022 Calib. No.: 00047
Certification File No.: 02116 Cert. Date: 12/09/2022 Cert. No.: 00038
Linearity File No.: 02117 Lin. Date: 12/09/2022 Lin. No.: 00038
Solution File No.: 02127 Soln. Date: 02/06/2023 Soln. No.: 00360
Sequential File No.: 02127 File Date: 02/06/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1004

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:03S		
Control 1 EC	0.102%	10:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:05S		
Control 2 EC	0.099%	10:05S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:06S		
Control 3 EC	0.100%	10:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:08S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 02/06/2023