

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	02115	Calib. Date:	12/09/2022
Certification File No.:	02116	Cert. Date:	12/09/2022
Linearity File No.:	02117	Lin. Date:	12/09/2022
Solution File No.:	02144	Soln. Date:	04/04/2023
Sequential File No.:	02144	File Date:	04/04/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0441
Solution Control Lot:	22160	Expires:	05/09/2024
		Bottle No.:	0891

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:27D		
Control 1 EC	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 2 EC	0.098%	10:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:30D		
Control 3 EC	0.099%	10:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 04/04/2023