

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 02115 Calib. Date: 12/09/2022 Calib. No.: 00047
Certification File No.: 02116 Cert. Date: 12/09/2022 Cert. No.: 00038
Linearity File No.: 02117 Lin. Date: 12/09/2022 Lin. No.: 00038
Solution File No.: 02146 Soln. Date: 05/03/2023 Soln. No.: 00364
Sequential File No.: 02146 File Date: 05/03/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 05/09/2024
Solution Control Lot: 22160 Bottle No.: 1013

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:27D		
Control 1 EC	0.101%	23:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:28D		
Control 2 EC	0.100%	23:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	23:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:29D		
Control 3 EC	0.100%	23:30D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	23:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/03/2023