

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 02115 **Calib. Date:** 12/09/2022 **Calib. No.:** 00047
Certification File No.: 02116 **Cert. Date:** 12/09/2022 **Cert. No.:** 00038
Linearity File No.: 02117 **Lin. Date:** 12/09/2022 **Lin. No.:** 00038
Solution File No.: 02145 **Soln. Date:** 05/03/2023 **Soln. No.:** 00363
Sequential File No.: 02145 **File Date:** 05/03/2023

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0441
Control Solution %: 0.100% **Expires:** 05/09/2024
Solution Control Lot: 22160 **Bottle No.:** 1156

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	17:09D		
Control 1 EC	----%	17:09D	22.2°C	SIMULATOR TEMP. ERROR
Control 1 IR	----%	17:09D	22.2°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	----%	--:--		
Control 2 EC	----%	--:--	N/A	
Control 2 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		
Control 3 EC	----%	--:--	N/A	
Control 3 IR	----%	--:--	N/A	
Ambient Air Blank	----%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/03/2023