

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 02115 **Calib. Date:** 12/09/2022 **Calib. No.:** 00047
Certification File No.: 02116 **Cert. Date:** 12/09/2022 **Cert. No.:** 00038
Linearity File No.: 02117 **Lin. Date:** 12/09/2022 **Lin. No.:** 00038
Solution File No.: 02155 **Soln. Date:** 05/28/2023 **Soln. No.:** 00365
Sequential File No.: 02155 **File Date:** 05/28/2023

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0441
Control Solution %: 0.100% **Expires:** 05/09/2024
Solution Control Lot: 22160 **Bottle No.:** 1154

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:29D		
Control 1 EC	0.096%	09:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		
Control 2 EC	0.097%	09:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	09:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		
Control 3 EC	0.096%	09:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	09:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:33D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 05/28/2023