

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	02156	Calib. Date:	05/31/2023
Certification File No.:	02157	Cert. No.:	00048
Linearity File No.:	02158	Cert. Date:	05/31/2023
Solution File No.:	02188	Lin. No.:	00039
Sequential File No.:	02188	Lin. Date:	05/31/2023
		Soln. Date:	08/05/2023
		Soln. No.:	00369
		File Date:	08/05/2023
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0440
Solution Control Lot:	23010	Expires:	01/11/2025
		Bottle No.:	1250

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:24D		
Control 1 EC	0.102%	12:25D	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:25D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		
Control 2 EC	0.101%	12:26D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:26D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		
Control 3 EC	0.101%	12:28D	34.1°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:28D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:28D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL      First Name: CHRISTOPHER      MI: J

Signature: \_\_\_\_\_



Badge No.: 128

Date: 08/05/2023