Calibrating Unit
New Standard Solution Report

Equipment Alcotest 7110 MKIII-C — Serial No.: ARWJ-0018

Location: SEASIDE HEIGHTS POLICE

Calibration File No.:	02156	Calib. Date:	05/31/2023	Calib. No.:	00048
Certification File No.:	02157	Cert. Date:	05/31/2023	Cert. No.:	00039
Linearity File No.:	02158	Lin. Date:	05/31/2023	Lin. No.:	00039
Solution File No.:	02199	Soln. Date:	08/21/2023	Soln. No.:	00370
Sequential File No.:	02199	File Date:	08/21/2023		

Calibrating Unit: WET — Model No.: CU-34 — Serial No.: DDWL S3-0440

Control Solution %: 0.100% — Expires: 05/09/2024

Solution Control Lot: 22160 — Bottle No.: 1427

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:40D		
Control 1 EC	0.101%	12:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:42D		
Control 2 EC	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:43D		
Control 3 EC	0.100%	12:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL — First Name: CHRISTOPHER — MI: J

Signature: _____ Badge No.: 128

Date: 08/21/2023