

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 02156 **Calib. Date:** 05/31/2023 **Calib. No.:** 00048
Certification File No.: 02157 **Cert. Date:** 05/31/2023 **Cert. No.:** 00039
Linearity File No.: 02158 **Lin. Date:** 05/31/2023 **Lin. No.:** 00039
Solution File No.: 02202 **Soln. Date:** 09/25/2023 **Soln. No.:** 00371
Sequential File No.: 02202 **File Date:** 09/25/2023

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0440
Control Solution %: 0.100% **Expires:** 05/09/2024
Solution Control Lot: 22160 **Bottle No.:** 1426

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:25D		
Control 1 EC	0.102%	10:26D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:26D		
Control 2 EC	0.100%	10:27D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:28D		
Control 3 EC	0.100%	10:28D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:29D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature:

Badge No.: 128

Date: 09/25/2023