

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	02203	Calib. Date: 09/29/2023
Certification File No.:	02204	Calib. No.: 00049
Linearity File No.:	02205	Cert. Date: 09/29/2023
Solution File No.:	02217	Cert. No.: 00040
Sequential File No.:	02217	Lin. Date: 09/29/2023
		Lin. No.: 00040
		Soln. Date: 11/22/2023
		Soln. No.: 00374
		File Date: 11/22/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	23330	Expires: 09/12/2025
		Bottle No.: 0775

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	20:45S		
Control 1 EC	0.098%	20:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:46S		
Control 2 EC	0.097%	20:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:48S		
Control 3 EC	0.097%	20:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	20:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL	First Name: CHRISTOPHER	MI: J
Signature: 		Badge No.: 128
		Date: 11/22/2023