

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01108	Calib. Date:	02/24/2015
Certification File No.:	01109	Cert. Date:	02/24/2015
Linearity File No.:	01110	Lin. Date:	02/24/2015
Solution File No.:	01132	Soln. Date:	04/25/2015
Sequential File No.:	01132	File Date:	04/25/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	06/20/2016
Solution Control Lot:	14F130			Bottle No.:	0752

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:13D		
Control 1 EC	0.098%	11:14D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 2 EC	0.098%	11:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16D		
Control 3 EC	0.100%	11:17D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:17D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: \_\_\_\_\_



Badge No.: 97517

Date: 04/25/2015