

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01108	Calib. Date: 02/24/2015
Certification File No.:	01109	Cert. No.: 00025
Linearity File No.:	01110	Cert. Date: 02/24/2015
Solution File No.:	01164	Lin. No.: 00018
Sequential File No.:	01164	Lin. Date: 02/24/2015
		Soln. No.: 00209
		Soln. Date: 06/29/2015
		File Date: 06/29/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWL S3-0440
Control Solution %:	0.100%		Expires: 06/20/2016
Solution Control Lot:	14F130		Bottle No.: 0762

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:01D		
Control 1 EC	0.097%	23:01D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:01D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:02D		
Control 2 EC	0.097%	23:03D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:03D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:03D		
Control 3 EC	0.097%	23:04D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:05D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 06/29/2015