

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01108	Calib. Date: 02/24/2015
Certification File No.:	01109	Cert. No.: 00025
Linearity File No.:	01110	Cert. No.: 00018
Solution File No.:	01172	Lin. No.: 00018
Sequential File No.:	01172	Soln. No.: 00210
		File Date: 07/11/2015

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDWL S3-0440
Control Solution %: 0.100%		Expires: 06/20/2016
Solution Control Lot: 14F130		Bottle No.: 0768

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:12D		
Control 1 EC	0.099%	03:13D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	03:13D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:14D		
Control 2 EC	0.098%	03:14D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	03:14D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:15D		
Control 3 EC	0.098%	03:16D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:16D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:16D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DIFABIO

First Name: LANCE

MI: S

Signature: _____



Badge No.: 3431

Date: 07/11/2015