

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01187	Calib. Date: 08/03/2015
Certification File No.:	01188	Cert. No.: 00026
Linearity File No.:	01189	Cert. No.: 00019
Solution File No.:	01212	Lin. No.: 00019
Sequential File No.:	01212	Soln. Date: 08/22/2015
		Soln. No.: 00217
		File Date: 08/22/2015

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 0767

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:36D		
Control 1 EC	0.101%	09:36D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:36D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:37D		
Control 2 EC	0.100%	09:38D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:38D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:38D		
Control 3 EC	0.100%	09:39D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:39D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:40D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 08/22/2015