

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01187	Calib. Date:	08/03/2015
Certification File No.:	01188	Cert. No.:	00026
Linearity File No.:	01189	Cert. Date:	08/03/2015
Solution File No.:	01216	Lin. No.:	00019
Sequential File No.:	01216	Soln. Date:	08/30/2015
		Soln. No.:	00218
		File Date:	08/30/2015

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	06/20/2016
Solution Control Lot:	14F130			Bottle No.:	0761

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:53D		
Control 1 EC	0.098%	13:54D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	13:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:54D		
Control 2 EC	0.097%	13:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	13:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:56D		
Control 3 EC	0.097%	13:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	13:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 08/30/2015