

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01187	Calib. Date: 08/03/2015
Certification File No.:	01188	Calib. No.: 00026
Linearity File No.:	01189	Cert. Date: 08/03/2015
Solution File No.:	01225	Cert. No.: 00019
Sequential File No.:	01225	Lin. Date: 08/03/2015
		Lin. No.: 00019
		Soln. Date: 09/06/2015
		Soln. No.: 00220
		File Date: 09/06/2015

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 0495

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:28D		
Control 1 EC	0.095%	09:29D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	09:29D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:30D		
Control 2 EC	0.095%	09:30D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	09:30D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:31D		
Control 3 EC	0.095%	09:32D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	09:32D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:32D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: ROEMMELE First Name: RICHARD MI: R
 Signature:  Badge No.: 97517
 Date: 09/06/2015