

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Seaside Heights Police Department 116 Sherman Ave. Seaside Heights, NJ 08751</b>		2. Contact: <b>Sgt. Richard Roemmele</b>	4. Date: <b>11/10/2015</b>
		3. Phone Number: <b>732-793-1800</b>	5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARWJ-0018</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Alcotest instrument returned from outside maintenance and placed back in service.**

10. Comments:  
**See Draeger Return and Repair form.**

- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☐ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☒ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgt. Steve Dellanoce #6027**

Name & Badge Number (Print)

Signature

*Sgt. Steve Dellanoce* #6027 11/10/15  
Date



# RETURN AND REPAIR FORM

# Dräger

## Customer Information

B:

S:

Company Name: Seaside Heights P.D., Seaside Heights, NJ

Date Received: 10/09/15

Date given to service: 10/09/15

Carrier: ☒ FedEx ☐ UPS ☐ USPS

Shipping Method: ☒ GRD ☐ 3DAY ☐ 2DAY  
☐ NDA-PRI ☐ NDA-STD

Product: ☐ 6510 ☐ 6810  
☐ 7110 ☐ 8610  
☐ 7510 ☐ DT5000  
☐ 7410 Upper-half ☐ 9510  
☐ 7410 Whole

Serial Number: AR WJ-0018

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: January 2016 (60725)

Description: ☐ A ☐ B ☐ Plus ☐ Demo ☐ Screener ☐ Trade In

## Accessories

☐ 110V A/C Adapter ☐ Regulator ☐ Printer Ribbon ☐ Printer Paper  
☐ Mouthpieces ☐ 9510 Stylus ☐ 9510 Top Cover ☐ Carrying Case  
☐ Dry Gas ☐ Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W
6808455	Fuel Cell	1	NC/W
6808486	Spacer Plates	2	NC/W

Repair Notes: Replaced defective fuel cell and spacer plates.

Fuel Cell = ARHE-1522

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/13/15



# Packing Slip

Customer no      Order No      Order date  
150060725      10809335      10/13/2015

Packing slip no.      Ship date  
80982935      10/13/2015

Please reference on inquiries

Ship to

150060729  
BOROUGH OF SEASIDE HEIGHTS POLICE  
ATTN: KELLY DWYER OR SGT. ROEMNELE  
OCEAN COUNTY  
116 SHERMAN AVENUE  
SEASIDE HEIGHTS, NJ 08751-2055  
USA

Payer

Your Purchase Order Number

KELLY DWYER OR SGT. ROEMNELE  
Any questions? Please contact:  
ROSEMARIE HUGHES

Customer

BOROUGH OF SEASIDE HEIGHTS  
MUNICIPAL BUILDING  
OCEAN COUNTY  
901 BLVD-PO BOX 38  
SEASIDE HEIGHTS, NJ 08751-0038  
USA

Your sales person

LINDA SALO  
Phone: 866-385-5900  
Fax: 972-929-1260

Delivery terms

FA

FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARWJ-0018 REC'D 10/9/15			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA			

**Draeger Safety Diagnostics Inc.**  
Accounting Address:  
101 Technology Drive  
Pittsburgh, PA 15275  
Tel: (412) 788-5537  
Fax: (412) 788-5598

**Remit Wire Transfers To:**  
Citizens Bank  
Acct. # 6209426615  
Acct. Name: Draeger Safety Diagnostics Inc.  
Transit Number: 036076150  
Federal ID Number: 84-1600159

**REMIT TO:**  
Draeger Safety Diagnostics Inc.  
P.O. Box 536410  
Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150060725    10809335    10/13/2015

Packing slip no.    Ship date  
80982935    10/13/2015  
Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description				
0030	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA				
0040	2 EA	6808486 PLATES, 7110--9510 Ordered / Back ordered : 0 / 0 EA				



STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Seaside Heights Police Department 116 Sherman Ave. Seaside Heights, NJ 08751</b>		2. Contact: <b>Sgt. Richard Roemmele</b>	4. Date: <b>09/16/2015</b>
		3. Phone Number: <b>732-793-1800</b>	5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARWJ-0018</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Complaint received of instrument experiencing control test failures**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

**Please replace Fuel Cell.**

**Last Known Sequential File #: 01225**

- ☐ 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- ☒ 12. The above Instrument/Component is placed out of service pending further evaluation.
- ☐ 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Sgt. Steve Dellanoce #6027**

Name & Badge Number (Print)

Signature

**Sgt. ASD #6027** **9/16/15**

Date