

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01231	Calib. Date: 11/10/2015
Certification File No.:	01232	Calib. No.: 00028
Linearity File No.:	01233	Cert. Date: 11/10/2015
Solution File No.:	01241	Cert. No.: 00020
Sequential File No.:	01241	Lin. Date: 11/10/2015
		Lin. No.: 00020
		Soln. No.: 00223
		File Date: 12/11/2015

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWL S3-0441
Control Solution %:	0.100%		Expires: 06/13/2017
Solution Control Lot:	15F139		Bottle No.: 0719

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:49S		
Control 1 EC	0.100%	00:50S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	00:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:50S		
Control 2 EC	0.100%	00:51S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:51S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:51S		
Control 3 EC	0.100%	00:52S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	00:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: \_\_\_\_\_



Badge No.: 120

Date: 12/11/2015