

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01231	Calib. Date: 11/10/2015
Certification File No.:	01232	Calib. No.: 00028
Linearity File No.:	01233	Cert. Date: 11/10/2015
Solution File No.:	01247	Cert. No.: 00020
Sequential File No.:	01247	Lin. Date: 11/10/2015
		Lin. No.: 00020
		Soln. Date: 01/08/2016
		Soln. No.: 00224
		File Date: 01/08/2016
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0441
Solution Control Lot:	15F139	Expires: 06/13/2017
		Bottle No.: 0724

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:18S		
Control 1 EC	0.099%	23:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	23:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:19S		
Control 2 EC	0.098%	23:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	23:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:21S		
Control 3 EC	0.099%	23:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	23:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____

Shawn Heckler #120

Badge No.: 120

Date: 01/08/2016