

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01261 Calib. Date: 03/01/2016 Calib. No.: 00029
Certification File No.: 01262 Cert. Date: 03/01/2016 Cert. No.: 00021
Linearity File No.: 01263 Lin. Date: 03/01/2016 Lin. No.: 00021
Solution File No.: 01281 Soln. Date: 03/30/2016 Soln. No.: 00227
Sequential File No.: 01281 File Date: 03/30/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 06/13/2017
Solution Control Lot: 15F139 Bottle No.: 0499

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:09S		
Control 1 EC	0.096%	22:10S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:10S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:11S		
Control 2 EC	0.095%	22:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	22:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:12S		
Control 3 EC	0.096%	22:13S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	22:13S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____

Shawn Heckler #120

Badge No.: 120

Date: 03/30/2016