

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01261	Calib. Date: 03/01/2016
Certification File No.:	01262	Cert. No.: 00029
Linearity File No.:	01263	Cert. Date: 03/01/2016
Solution File No.:	01327	Lin. No.: 00021
Sequential File No.:	01327	Soln. No.: 00237
		File Date: 07/30/2016

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDWL S3-0441
Control Solution %:	0.100%		Expires: 05/17/2018
Solution Control Lot:	201605D		Bottle No.: 0237

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:33D		
Control 1 EC	0.102%	22:33D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	22:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:34D		
Control 2 EC	0.102%	22:35D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	22:35D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:35D		
Control 3 EC	0.101%	22:36D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	22:36D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:37D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: ROEMMEL

First Name: RICHARD

MI: R

Signature: _____



Badge No.: 97517

Date: 07/30/2016