

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01348 Calib. Date: 10/28/2016 Calib. No.: 00032
Certification File No.: 01349 Cert. Date: 10/28/2016 Cert. No.: 00023
Linearity File No.: 01350 Lin. Date: 10/28/2016 Lin. No.: 00023
Solution File No.: 01364 Soln. Date: 11/24/2016 Soln. No.: 00240
Sequential File No.: 01364 File Date: 11/24/2016

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440
Control Solution %: 0.100% Expires: 05/17/2018
Solution Control Lot: 201605D Bottle No.: 0238

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:52S		
Control 1 EC	0.098%	22:53S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	22:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:53S		
Control 2 EC	0.098%	22:54S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	22:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:55S		
Control 3 EC	0.098%	22:55S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:56S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____



Badge No.: 120

Date: 11/24/2016