

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01348 **Calib. Date:** 10/28/2016 **Calib. No.:** 00032
Certification File No.: 01349 **Cert. Date:** 10/28/2016 **Cert. No.:** 00023
Linearity File No.: 01350 **Lin. Date:** 10/28/2016 **Lin. No.:** 00023
Solution File No.: 01376 **Soln. Date:** 01/16/2017 **Soln. No.:** 00242
Sequential File No.: 01376 **File Date:** 01/16/2017

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDWL S3-0440
Control Solution %: 0.100% **Expires:** 05/17/2018
Solution Control Lot: 201605D **Bottle No.:** 0242

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:02S		
Control 1 EC	0.096%	21:03S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	21:03S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:03S		
Control 2 EC	0.096%	21:04S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	21:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:05S		
Control 3 EC	0.096%	21:06S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.096%	21:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:06S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: _____

Badge No.: 97517

Date: 01/16/2017