

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
<b>Location:</b>	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01348	Calib. Date:	10/28/2016
Certification File No.:	01349	Cert. Date:	10/28/2016
Linearity File No.:	01350	Lin. Date:	10/28/2016
Solution File No.:	01387	Soln. Date:	02/11/2017
Sequential File No.:	01387	File Date:	02/11/2017
		Calib. No.:	00032
		Cert. No.:	00023
		Lin. No.:	00023
		Soln. No.:	00243

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWL S3-0440
Control Solution %:	0.100%			Expires:	06/07/2018
Solution Control Lot:	16140			Bottle No.:	1328

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	08:43S		
Control 1 EC	0.096%	08:44S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.096%	08:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:45S		
Control 2 EC	0.096%	08:45S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:46S		
Control 3 EC	0.096%	08:47S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	08:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:47S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: \_\_\_\_\_



Badge No.: 120

Date: 02/11/2017