

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01348 Calib. Date: 10/28/2016 Calib. No.: 00032
Certification File No.: 01349 Cert. Date: 10/28/2016 Cert. No.: 00023
Linearity File No.: 01350 Lin. Date: 10/28/2016 Lin. No.: 00023
Solution File No.: 01392 Soln. Date: 02/17/2017 Soln. No.: 00244
Sequential File No.: 01392 File Date: 02/17/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440
Control Solution %: 0.100% Expires: 06/07/2018
Solution Control Lot: 16140 Bottle No.: 1339

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:20S		
Control 1 EC	0.096%	19:20S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	19:20S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:21S		
Control 2 EC	0.095%	19:22S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	19:22S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:22S		
Control 3 EC	0.096%	19:23S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.095%	19:23S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 02/17/2017