

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	Calib. No.:	00032
Calibration File No.:	01348	Cert. Date:	10/28/2016
Certification File No.:	01349	Lin. Date:	10/28/2016
Linearity File No.:	01350	Soln. Date:	03/17/2017
Solution File No.:	01411	File Date:	03/17/2017
Sequential File No.:	01411		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0440
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1032

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:33S		
Control 1 EC	0.097%	23:33S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	23:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:34S		
Control 2 EC	0.096%	23:35S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	23:35S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:35S		
Control 3 EC	0.096%	23:36S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	23:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:37S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: 

Badge No.: 120

Date: 03/17/2017