

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01348	Calib. Date: 10/28/2016
Certification File No.:	01349	Calib. No.: 00032
Linearity File No.:	01350	Cert. Date: 10/28/2016
Solution File No.:	01420	Cert. No.: 00023
Sequential File No.:	01420	Lin. Date: 10/28/2016
		Lin. No.: 00023
		Soln. Date: 03/31/2017
		Soln. No.: 00250
		File Date: 03/31/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1036

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	07:40S		
Control 1 EC	0.098%	07:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	07:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:42S		
Control 2 EC	0.097%	07:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43S		
Control 3 EC	0.098%	07:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____

Badge No.: 120

Date: 03/31/2017