

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE		
Calibration File No.:	01427	Calib. Date:	04/21/2017
Certification File No.:	01428	Cert. Date:	04/21/2017
Linearity File No.:	01429	Lin. Date:	04/21/2017
Solution File No.:	01442	Soln. Date:	05/14/2017
Sequential File No.:	01442	File Date:	05/14/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWL S3-0441
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1241

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:46D		
Control 1 EC	0.098%	20:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:47D		
Control 2 EC	0.098%	20:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:49D		
Control 3 EC	0.097%	20:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	20:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: \_\_\_\_\_



Badge No.: 97517

Date: 05/14/2017