

**ALCOHOL INFLUENCE REPORT FORM, ALCOTEST 7110 MKIII-C
SEASIDE HEIGHTS POLICE**

Department Case No.: 17-5303
Summons No(s): SH-155164
Sequential File No.: 01455

Subject

Last Name: LEONE First Name: MICHAEL MI: A
D.O.B.: 08/01/1994 Age: 22 Gender: MALE Ht: 5 ft. 09 in. Wt: 165 lbs.
Driver License Number: L26085446108942 Issuing State: NJ

Arresting Officer

Last Name: VARADY First Name: SEAN MI: P
Badge No.: 0460 Arrest Date: 06/01/2017 Arrest Time: 02:49D Arrest Location: 1526

Instrument

Alcotest 7110 MKIII-C Serial No.: ARWJ-0018
Location: SEASIDE HEIGHTS POLICE
Calibration File No.: 01427 Calib. Date: 04/21/2017 Calib. No.: 00033
Certification File No.: 01428 Cert. Date: 04/21/2017 Cert. No.: 00024
Linearity File No.: 01429 Lin. Date: 04/21/2017 Lin. No.: 00024
Solution File No.: 01442 Soln. Date: 05/14/2017 Soln. No.: 00253
Sequential File No.: 01455 File Date: 06/01/2017


Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441
Control Solution %: 0.100% Expires: 12/02/2018
Solution Control Lot: 16390 Bottle No.: 1241

Breath Test Information

| | | | | | | Date of Test: 06/01/2017 |
|-------------------|--------|--------|--------|----------|-----------|--------------------------|
| Function | Result | Time | Volume | Duration | Temp. | Error Message |
| | %BAC | HH:MM | (L) | Sec (s) | Sim. (°C) | |
| Ambient Air Blank | 0.000% | 03:37D | | | | |
| Control Test 1 | | | | | 34.0°C | |
| EC Result | 0.097% | 03:38D | | | | |
| IR Result | 0.096% | 03:38D | | | | |
| Ambient Air Blank | 0.000% | 03:39D | | | | |
| Breath Test 1 | | | --L | --s | | SUBJECT REFUSED |
| EC Result | ----% | --:-- | | | | |
| IR Result | ----% | --:-- | | | | |
| Ambient Air Blank | 0.000% | 03:40D | | | | |
| Control Test 2 | | | | | 34.0°C | |
| EC Result | 0.096% | 03:40D | | | | |
| IR Result | 0.096% | 03:40D | | | | |
| Ambient Air Blank | 0.000% | 03:41D | | | | |

TEST RESULT: SUBJECT REFUSED

Breath Test Operator

Last Name: VARADY First Name: SEAN MI: P
Signature:  Badge No.: 0460
Date: 06/01/2017

Copy Given to Subject