

# Calibrating Unit

## New Standard Solution Report

**Equipment**

Alcotest 7110 MKIII-C  
Location: SEASIDE HEIGHTS POLICE  
Serial No.: ARWJ-0018  
Calibration File No.: 01427  
Calib. Date: 04/21/2017  
Calib. No.: 00033  
Certification File No.: 01428  
Cert. Date: 04/21/2017  
Cert. No.: 00024  
Linearity File No.: 01429  
Lin. Date: 04/21/2017  
Lin. No.: 00024  
Solution File No.: 01457  
Soln. Date: 06/04/2017  
Soln. No.: 00254  
Sequential File No.: 01457  
File Date: 06/04/2017

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16390  
Model No.: CU-34  
Serial No.: DDWL S3-0441  
Expires: 12/02/2018  
Bottle No.: 1196

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	02:51D		
Control 1 EC	0.098%	02:52D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	02:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:52D		
Control 2 EC	0.098%	02:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	02:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:54D		
Control 3 EC	0.098%	02:55D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	02:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: \_\_\_\_\_



Badge No.: 97517

Date: 06/04/2017