

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01427 Calib. Date: 04/21/2017 Calib. No.: 00033  
Certification File No.: 01428 Cert. Date: 04/21/2017 Cert. No.: 00024  
Linearity File No.: 01429 Lin. Date: 04/21/2017 Lin. No.: 00024  
Solution File No.: 01478 Soln. Date: 07/01/2017 Soln. No.: 00256  
Sequential File No.: 01478 File Date: 07/01/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0441  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1242

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:30D		
Control 1 EC	0.099%	14:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:31D		
Control 2 EC	0.098%	14:32D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:32D		
Control 3 EC	0.098%	14:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: 

Badge No.: 128

Date: 07/01/2017