

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01427	Calib. Date: 04/21/2017
Certification File No.:	01428	Calib. No.: 00033
Linearity File No.:	01429	Cert. Date: 04/21/2017
Solution File No.:	01500	Cert. No.: 00024
Sequential File No.:	01500	Lin. Date: 04/21/2017
		Lin. No.: 00024
		Soln. Date: 07/30/2017
		Soln. No.: 00258
		File Date: 07/30/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0441
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1400

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:52D		
Control 1 EC	0.098%	15:53D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:54D		
Control 2 EC	0.097%	15:54D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	15:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:55D		
Control 3 EC	0.096%	15:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	15:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: _____

Badge No.: 128

Date: 07/30/2017

