

Calibrating Unit

New Standard Solution Report

| | | | |
|-------------------------|------------------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWJ-0018 |
| Location: | SEASIDE HEIGHTS POLICE | Calib. No.: | 00034 |
| Calibration File No.: | 01531 | Cert. Date: | 10/12/2017 |
| Certification File No.: | 01532 | Lin. Date: | 10/12/2017 |
| Linearity File No.: | 01533 | Soln. Date: | 11/11/2017 |
| Solution File No.: | 01541 | File Date: | 11/11/2017 |
| Sequential File No.: | 01541 | | |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWL S3-0440 |
| Solution Control Lot: | 16390 | Expires: | 12/02/2018 |
| | | Bottle No.: | 1030 |

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 02:34S | | |
| Control 1 EC | 0.097% | 02:35S | 33.9°C | *** TEST PASSED *** |
| Control 1 IR | 0.097% | 02:35S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:36S | | |
| Control 2 EC | 0.097% | 02:36S | 33.9°C | *** TEST PASSED *** |
| Control 2 IR | 0.095% | 02:36S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 02:37S | | |
| Control 3 EC | 0.095% | 02:38S | 33.9°C | CONTROL TEST FAILED |
| Control 3 IR | 0.092% | 02:38S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 02:38S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: _____



Badge No.: 120

Date: 11/11/2017