

# Calibrating Unit

## New Standard Solution Report


<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01531	Calib. Date: 10/12/2017
Certification File No.:	01532	Calib. No.: 00034
Linearity File No.:	01533	Cert. Date: 10/12/2017
Solution File No.:	01542	Cert. No.: 00025
Sequential File No.:	01542	Lin. Date: 10/12/2017
		Lin. No.: 00025
		Soln. Date: 11/11/2017
		Soln. No.: 00263
		File Date: 11/11/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1240

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:53S		
Control 1 EC	0.097%	05:54S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	05:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:54S		
Control 2 EC	0.096%	05:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	05:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:56S		
Control 3 EC	0.096%	05:56S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	05:56S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:57S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: HECKLER	First Name: SHAWN	MI: M
Signature: 	Badge No.: 120	Date: 11/11/2017