

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01531 Calib. Date: 10/12/2017 Calib. No.: 00034  
Certification File No.: 01532 Cert. Date: 10/12/2017 Cert. No.: 00025  
Linearity File No.: 01533 Lin. Date: 10/12/2017 Lin. No.: 00025  
Solution File No.: 01547 Soln. Date: 11/12/2017 Soln. No.: 00264  
Sequential File No.: 01547 File Date: 11/12/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1045

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:59S		
Control 1 EC	0.097%	21:59S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:59S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:00S		
Control 2 EC	0.095%	22:01S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:01S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:02S		
Control 3 EC	0.096%	22:02S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:02S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:03S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: LINNELL

First Name: CHRISTOPHER

MI: J

Signature: \_\_\_\_\_

Badge No.: 128

Date: 11/12/2017