

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWJ-0018
Location:	SEASIDE HEIGHTS POLICE	
Calibration File No.:	01531	Calib. Date: 10/12/2017
Certification File No.:	01532	Calib. No.: 00034
Linearity File No.:	01533	Cert. Date: 10/12/2017
Solution File No.:	01554	Cert. No.: 00025
Sequential File No.:	01554	Lin. Date: 10/12/2017
		Lin. No.: 00025
		Soln. Date: 12/12/2017
		Soln. No.: 00265
		File Date: 12/12/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWL S3-0440
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1201

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	23:11S		
Control 1 EC	0.098%	23:11S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	23:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:12S		
Control 2 EC	0.096%	23:13S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	23:13S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:13S		
Control 3 EC	0.097%	23:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	23:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: ROEMMELE

First Name: RICHARD

MI: R

Signature: \_\_\_\_\_



Badge No.: 7517

Date: 12/12/2017