

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWJ-0018  
Location: SEASIDE HEIGHTS POLICE  
Calibration File No.: 01531 Calib. Date: 10/12/2017 Calib. No.: 00034  
Certification File No.: 01532 Cert. Date: 10/12/2017 Cert. No.: 00025  
Linearity File No.: 01533 Lin. Date: 10/12/2017 Lin. No.: 00025  
Solution File No.: 01561 Soln. Date: 12/29/2017 Soln. No.: 00266  
Sequential File No.: 01561 File Date: 12/29/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWL S3-0440  
Control Solution %: 0.100% Expires: 12/02/2018  
Solution Control Lot: 16390 Bottle No.: 1145

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:30S		
Control 1 EC	0.097%	04:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	04:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:31S		
Control 2 EC	0.095%	04:32S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	04:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:33S		
Control 3 EC	0.096%	04:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	04:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:34S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: HECKLER

First Name: SHAWN

MI: M

Signature: \_\_\_\_\_

*Shawn Heckler #120*

Badge No.: 120

Date: 12/29/2017