

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|------------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARWJ-0018 |
| Location: | SEASIDE HEIGHTS POLICE | | |
| Calibration File No.: | 01531 | Calib. Date: | 10/12/2017 |
| Certification File No.: | 01532 | Cert. Date: | 10/12/2017 |
| Linearity File No.: | 01533 | Lin. Date: | 10/12/2017 |
| Solution File No.: | 01564 | Soln. Date: | 01/01/2018 |
| Sequential File No.: | 01564 | File Date: | 01/01/2018 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWL S3-0440 |
| Solution Control Lot: | 16390 | Expires: | 12/02/2018 |
| | | Bottle No.: | 1049 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 00:46S | | |
| Control 1 EC | 0.095% | 00:47S | 33.8°C | *** TEST PASSED *** |
| Control 1 IR | 0.095% | 00:47S | 33.8°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 00:47S | | |
| Control 2 EC | 0.094% | 00:48S | 33.9°C | CONTROL TEST FAILED |
| Control 2 IR | 0.096% | 00:48S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 00:49S | | |
| Control 3 EC | 0.094% | 00:50S | 33.9°C | CONTROL TEST FAILED |
| Control 3 IR | 0.095% | 00:50S | 33.9°C | CONTROL TEST FAILED |
| Ambient Air Blank | 0.000% | 00:50S | | |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: LINNELL

First Name: CHRISTOPHER

MI: -

Signature: _____

Badge No.: 128

Date: 01/01/2018

